

PROVIDE WRITTEN CORRESPONDENCE OR REPORT OF CALIBRATION STATING AND INCLUDING THE FOLLOWING FOR EACH INSTRUMENT:

- RESPONSIBLE SIGNATURE(S), LOCATION, DATE OF CALIBRATION, AND IDENTIFICATION OF THE ACTIVITY THAT PERFORMED CALIBRATION

- ACCURACY AND TRACEABILITY TO NIST AND APPLICABLE NIST TEST NUMBERS FOR THE EQUIPMENT USED TO PERFORM CALIBRATION

- ENVIRONMENTAL CONDITIONS UNDER WHICH CALIBRATION WAS PERFORMED

- ACTUAL TEST POINT READINGS UP SCALE AND DOWN SCALE FOR A MINIMUM OF FIVE (5) EQUALLY SPACED TEST POINTS, SUBSTANTIATING THAT SUBJECT INSTRUMENT MEETS MANUFACTURER'S ADVERTISED ACCURACY

THE VENDOR'S CALIBRATION SYSTEM, SUBJECT TO AUDIT, MUST MEET THE REQUIREMENTS OF MIL-STD-45662, ANSI-Z540-1 OR ISO 10012-1, AS A MINIMUM. THE VENDOR'S WRITTEN CORRESPONDENCE, SHOWING THE ABOVE DATA, MUST BE PROVIDED TO THIS ACTIVITY.