

CONTINUOUSLY MONITORED 24 HOUR BURN-IN TEST

- 1) ONE COPY OF TEST RESULTS (SAVED IN MICROSOFT EXCEL) FOR A 24-HOUR CONTINUOUSLY MONITORED BURN-IN TEST ARE REQUIRED FOR EACH PRODUCTION UNIT.
- 2) TESTS SHALL BE CONDUCTED AT SELLER'S FACILITIES AS THE FINAL STEP IN THE TEST SEQUENCE. BURN-IN RESULTS PERFORMED BY THE MANUFACTURER WILL NO LONGER BE ACCEPTED.
- 3) TESTS SHALL BE CONDUCTED AT AMBIENT TEMPERATURE AND PRESSURE.
- 4) BURN-IN TEST RESULTS SHALL BE SAVED AS A MICROSOFT EXCEL FILE AND ADHERE TO THE FOLLOWING FORMAT:
 - A MINIMUM OF ONE SAMPLE PER SECOND
 - ONE FILE PER UNIT
 - ONE TAB PER FILE

COLUMN-A: SHALL INDICATE THE DATE IN MM-DD-YY FORMAT.
COLUMN-B: SHALL INDICATE THE TIME IN HH.MM.SS.00 FORMAT.
COLUMN-C: SHALL INDICATE THE PRESSURE (PSIA) IN 00.000* FORMAT.
COLUMN-D: SHALL INDICATE THE CURRENT OUTPUT (mA) IN 0.000* FORMAT.

NOTE*: COLUMNS C AND D SHALL BOTH DISPLAY 3 DIGITS TO RIGHT OF DECIMAL.
- 5) EACH FILE SHALL ADHERE TO THE FOLLOWING NAMING CONVENTION:
VPAR XXXXXX-XXX EB-XXXX-X BURN-IN DATA FOR SN XXXXXX MM-DD-YY.XLSX
- 6) EVIDENCE OF ALL TESTS AND INSPECTIONS PERFORMED AS REQUIRED BY THE SPECIFICATIONS MUST BE MAINTAINED BY THE SELLER.
- 7) PROVIDE A CERTIFICATION OF CONFORMANCE TO THE BURN-IN TEST REQUIREMENTS. THE CERTIFICATION SHALL BE IN THE FOLLOWING FORMAT:
"I CERTIFY THAT THE ATTACHED EXCEL FILES CONTAINS THE TEST RESULTS PERFORMED IN ACCORDANCE WITH SC-76-9CB FOR ___ LIST ALL FILE NAMES ___
."
- 8) TEST REPORTS AND/OR OTHER VERIFICATION OF CONFORMANCE DOCUMENTS SHALL BE SIGNED AND DATED BY THE SELLER'S AUTHORIZED REPRESENTATIVE.