

Coded Note Number: **C3020**

Revision Level: **C001**

Date: **March 21, 2017**

Title: **MEASURING EQUIPMENT CALIBRATION CERTIFICATION**

This revision history is provided for convenience and does not alleviate the supplier's responsibility with understanding and complying with the full coded note.

Change from Revision C – Editorial change to reformat to current standard.

Bolded font indicates changed/added content.

[Text deleted] inserted in the document indicates the removal of content.

Quantity: One (1) copy.

Certificate of calibration and supporting test report/data sheet shall be supplied that includes:

1. Name of organization performing the calibration.
2. Identification of the equipment calibrated.
3. Date of the calibration.
4. Accuracy as found (specific data required if out of tolerance).
5. Accuracy as returned (specific data or reference to applicable manufacturer's specifications required).
6. Calibration sources' (standards) identification and calibration status.
7. Supporting test report/data sheet identification.
8. Pertinent environmental conditions of the calibration.
9. Purchase order number.
10. A statement that the standards used for the calibration are traceable to American National Standards, and that all standards and calibration processes employed are in accordance with one of the following: ANSI/NCSL Z540-1, ISO/IEC 17025, ISO 10012, ISO GUIDE 25, OR NAVAIR 17-35QAC-01.

The certificate shall be signed by an authorized representative.