

Coded Note Number: **LI018**

Revision: **C002**

Date: **February 9, 2016**

Title: **Calibration Results**

This revision history is provided for convenience and does not alleviate the supplier's responsibility with understanding and complying with the full coded note.

Bolded font indicates changed/added content.

[Text deleted] inserted in the document indicates the removal of content.

For 0.1% accurate test instruments select 10 equally spaced points at or near the cardinal points (major graduations on the dial usually labeled with a pressure value) as allowed by the dial graduations.

Perform pressure gauge measurements and record values in increasing order for the selected points.

Except for the highest value, perform pressure gauge measurements and record values in decreasing order for the selected points.

Supplier is to record the data obtained above and provide 2 copies of the calibration results with the material.